

body

# cellulite fat reduction

November-January 2005  
Australian Cosmetic Surgery - pg 100  
VelaSmooth - "Cellulite & Fat Reduction"

**COSMETIC**

**SURGERY MAGAZINE**

RRP \$10.95 incl. GST | NZ \$11.95 incl. GST  
ISSUE 26 • NOV - JAN 2005

INCORPORATING  
BODY HEALTH & BEAUTY

LIGHT  
THERAPY  
FOR ACNE

SILICONE  
IMPLANTS:  
USA UPDATE

GLOBAL TRENDS  
IN COSMETIC  
MEDICINE



## Bipolar radiofrequency, infrared heat and pulsatile suction in the non-surgical treatment of focal lipodystrophy and cellulite.

R Stephen Mulholland, MD, Toronto, Canada

Presented at the Australasian College of Cosmetic Surgery Conference, August 27, 2004, Canberra, Australia.

### Abstract

15 Patients and 66 anatomical sites were treated with a bipolar radiofrequency, infrared heat and pulsatile suction device (VelaSmooth, Syneron Medical Ltd, Israel). The protocol consisted of 16 treatments over 8 weeks (two treatments per week). Each treatment consisted of 3-6 passes over focal areas of lipodystrophy and cellulite. The end point for each treatment zone was warmth and erythema. Each treatment required 45 minutes. Weekly measurements included circumferential thickness of the thighs, hips and abdomen, each measured standard distances from stable bony landmarks. Weekly body weight was measured and metabolic indices were recorded monthly, including a complete blood count, electrolytes, cholesterol, triglycerides, VLDL, HDL and LDL as well as liver enzymes and a liver and kidney ultrasound.

There were 15 study patients, all female, with an average age of 46 years. The average pre-treatment weight was 76.73 kg and the post-treatments weight 76.64 kg. Mean circumferential reductions of the thigh of 4.18cm, abdomen 5.73 cm and hips 5.86 cm were achieved.

**'Cellulite is estimated to affect over 50% of women over 50 years of age and can affect women with focal lipodystrophy and also very thin women with no areas of excess fat'**

Patient-based assessment of cellulite and skin texture improvement was 62% (range 20-80%) and blinded doctor based evaluation was 40%. There were no complications in the study and overall patient happiness was over 80%. We conclude that the VelaSmooth, bipolar radiofrequency, infrared energy and vacuum suction device can achieve predictable, safe and clinically non-surgical reduction of focal lipodystrophic regions, smoothing of cellulitic skin and skin firming.

Liposuction is the most commonly performed invasive cosmetic surgery procedure in the United States. Liposuction surgery has traditionally boasted high patient happiness scores, but still carries with it significant peri-operative morbidity including death or medical illness, pulmonary emboli, scarring, seroma, skin irregularity and dysaesthesia. Cellulite is estimated to affect over 50% of women over 50 years of age and can affect women with focal lipodystrophy and also very

thin women with no areas of excess fat.

The demand for non-surgical and minimally invasive cosmetic enhancement procedures has witnessed the rapid growth of products and procedures such as Botox, soft tissue fillers, peels, laser hair removal, fotofacial and photorejuvenation, microdermabrasion, thread and suture based face-lifting. In the past, minimally invasive body contouring and cellulite improvement services have included endermology(3), mesotherapy(4) and carboxytherapy(5), with only modest and generally inconsistent fat contouring or cellulite outcomes. There would be significant demand and interest in a non-surgical procedure that delivered 'liposuction-like' contour reductions of focal lipodystrophies as well as significant smoothing of cellulite.

### Materials & methods

15 female patients were selected for study. Inclusion criteria were areas of focal fat accumulation and cellulite, an understanding of the potential risks and benefits, were medically well and not pregnant. The average age was 46 years old. All patients underwent a standardised protocol of twice weekly treatments with the VelaSmooth device. The VelaSmooth delivers bipolar radiofrequency energy, infrared heat energy and pulsatile vacuum suction through a large hand-held applicator that is administered directly to the skin. Each treatment lasted approximately 30-45 minutes and each zone was treated with 3-6 passes of the VelaSmooth applicator with the end point being significant erythema and warmth radiating from the treated skin.

The VelaSmooth applicator has two larger rollers which act as large RF electrodes with an IR lamp placed above and between the two electrodes. The clinical treatment is performed on each area of fat and/or cellulite twice weekly for an average of eight weeks. The clinical treatment requires some finesse and the end point for each zone is erythema and warmth

64 zones were treated in the 15 patients. All patients had their inner and outer thighs, abdomen, waist and hips treated for lipodystrophy, and the back and front of thighs for cellulite. There were no special dietary, exercise or water consumption instructions to participate in the study. Patients had weekly circumferential measurements taken of their thighs, waist-abdomen and hips as well as body weight. The circumferential measurements were always conducted at a specific and consistent distance from an anatomical bony landmark, the ASIS for the hips, the umbilicus for the abdomen-waist and an inferior distance from the inguinal tubercle and greater trochanter for the thighs. Pre-study and post-study analysis were made of

AUSTRALIAN COSMETIC SURGERY

November-January 2005  
Australian Cosmetic Surgery - pg 100  
VelaSmooth - "Cellulite & Fat Reduction"



## body

the following indices: complete blood count, electrolytes, liver enzymes, renal function, cholesterol, triglyceride and lipoprotein levels and liver and kidney ultrasounds. Pre and post-study digital photographs were taken of the anatomical lipodystrophic regions as well as areas of cellulite. Following the study patients were asked to grade their cellulite improvement on a linear analogue scale (0-100%) and the before and after photos of the cellulite underwent independent evaluation by a blinded dermatologist, using a standardised cellulite score(6).

### Results

All study patients completed the eight week, 16 treatment study. The average circumferential reductions were 4.18cm for the outer-inner thighs (range 1.5-6.0cm), 5.73cm for the abdomen-waist (range 3-8.0cm) and 5.86 cm for the hips (range 3-9.0cm). Table 1 details all the circumferential lipodystrophy data.

**TABLE 1** Pre-study and post-study circumferential reduction measurements.

Number of patients = 15	Pre-study average	Post-study average	Average change
Weight	76.73kg	76.64kg	0.09kg
Thigh Circumference (n=15)	65.50cm	61.32cm	4.18cm
Abdomen-Waist Circumference (n=15)	87.64cm	81.91cm	5.73cm
Hip Circumference (n=15)	107.64cm	101.77cm	5.86cm
Braline (n=2)	95.4cm	91.0cm	4.4cm
Calves (n=2, 1 patient)	34.6cm	32.2cm	2.2cm
Upper Arms (n=2, 1 patient)	18.8cm	17.4cm	1.4cm

All patients underwent a metabolic profiling before and after the study and despite altering significantly the fat depositions, there were no significant alterations in the metabolic indices as summarised in Table 2.

**TABLE 2** Pre and post-study metabolic indices and profiles.

	Pre-study	Post-study
Hb	134	133
Urea	3.0	2.9
Creatinine	70	72
Na+	139	140
K+	3.4	3.2
Cl-	110	109
Cholesterol	3.2	3.1
Triglycerides	1.35	1.25
VLDL, HDL/LDL	normal	normal
AST	28	25
Alkaline Phosphatase	76	70
Liver Ultrasound	Normal	Normal
Kidney Ultrasound	Normal	Normal

102

Patients were asked to complete patient perception linear analog score grading their perceived cellulite improvement. The average score was 62%, with a range of 20-80%. Pre and post-study digital photographs were analyzed by a blinded dermatologist who applied a standard cellulite grading score (6) to the photographs and the unblinded analysis revealed an average 40% improvement in the cellulite.

Patients were also asked to record their level of happiness with their fat contouring and cellulite smoothing results. 81% of study patients were very happy or happy with their results, 13% were satisfied and 6% were unhappy. 90% of patients would recommend the program to their friends. All patients made subjective comments that their skin was 'tighter' or 'firmer' after the program and several commented on improved stretch marks over the abdomen, but these outcome variables were not objectified by the study design.

### Discussion

The tremendous growth in interest in aesthetic plastic surgery has been fuelled in large part by minimal and non-invasive procedures. Most of our non-invasive enhancements have been aimed at rejuvenation of the face and neck. Mesotherapeutic chemical modulation and subcutaneous carbon dioxide insufflation (carboxytherapy) of localised fat and cellulite have been common in Europe, South America and Canada for some years, but the results in body contour alteration and cellulite improvement have been inconsistent and modest at best (3-5). Endermology provided early promise in the treatment of cellulite and body contouring, but the results and long-term success have not been overwhelming, which has limited the success of this modality in the offices of most plastic surgeons. By contrast the use of radiofrequency and infrared heat, together with pulsatile suction in the form of the VelaSmooth™ device provides a very consistent reduction in focal lipodystrophy, cellulite smoothing and skin tightening.

In this study, there was a 100% response in circumferential reduction in lipodystrophic regions. In the absence of weight loss or dietary manipulations one can conclude that the fat reduction contouring was a result of the energies delivered. The author postulates that the radiofrequency and infrared energy delivers a critical amount of thermal effect to the deep fat with each treatment. Patients describe their treated areas actually radiate heat for 1-2 hours after each treatment. This critical heating of the fat stimulates a lipolytic pathway that continues over the 8-week program to break down and metabolise the triglyceride contents of the localised Lipocyte into its component free fatty-acid, choline and glycerol. The body's normal transport mechanisms and lipoproteins take the broken down triglyceride back to the liver for metabolism, utilisation or excretion. During the

November-January 2005  
Australian Cosmetic Surgery - pg 100  
VelaSmooth - "Cellulite & Fat Reduction"



study, none of the lipoprotein, triglyceride, cholesterol or liver function tests changed, suggesting that the slow metabolic triglyceride consumption stimulated by the treatments appears to be safe.

Thermogenic RF mediated focal break down of fat has also been anecdotally reported using the Thermage

### 'A majority of patients with stretch marks reported improvement'

monopolar Thermcool system, however, this focal fat loss has been irregular and aesthetically displeasing. The thermogenic lipolytic contour improvements with the VelaSmooth proved to be very safe and the contour alterations smooth and consistent. With proper weight maintenance, patients' results might be expected to last for a long time. If, however, you add a caloric stimulus, the shrunken lipocyte will likely expand.

Over the first month of treatments, the fat reduction was just starting to be measurable, but it was not until after 8 weeks that the reductions had maximised and had started to plateau. Some larger patients may benefit from an additional month of treatment. There is definitely some technique and finesse involved in the treatments, particularly in achieving an endpoint.

Cellulite is a poorly understood disease, likely of chronic lymphatic and microcirculatory compromise, lipolymphedema, fibrosis and contraction and, ultimately, the clinical manifestation of cellulite. The Bipolar RF, IR and suction treatments showed noticeable and consistent cellulite improvements. The mechanism is possibly related to mechanical and thermal enhanced blood flow, lymphatic drainage and an increase in metabolic oxygenation to the cellulite tissue with a resultant decompression of the lipolymphedema and reduced sclerosis and fibrosis. However, unlike the fat contouring, which may be long-lasting, the improvements in cellulite will need a maintenance treatment every 4-6 weeks.

The overall patient observation of increased skin tightness may be related to the dermal matrix heating observed with bipolar RF and skin tightening noticed on

the face and upper torso (7). A majority of patients with stretch marks, even mature hypertrophic striatum, reported improvement.

The use of this device employed by the author has other potential benefits, but is not a formal part of this study, specifically after liposuction. The device and twice weekly protocol is instituted around 8-12 weeks postoperatively and the author's incidence of post-liposuction revision has decreased by approximately 50% and the cellulite or skin texture irregularity incidence has also significantly decreased.


#### CONCLUSION

- 1 RF-IR and pulsatile suction provide consistent and pleasing fat contour results in areas of focal lipodystrophy without a knife, surgery or recovery.
- 2 Cellulite and skin tone also improves.
- 3 Overall, the treatment of focal lipodystrophy and cellulite using RF-IR and suction, has a high patient acceptance, low incidence of complication and a high patient satisfaction score.
- 4 Bipolar RF, IR and suction may help reduce post-liposuction contour irregularities and lowered revision rates. **acsm**

*References available upon request.*



**Figure 1 Patient pre-treatment (left) and post-treatment (right). Hip and abdomen lipodystrophy circumferential reduction of 6.5cm reduction on the hips of 5.5cm**



**Figure 2 Pre-treatment (left) and post treatment (right), with 9cm reduction of abdomen to waist and 9cm from the hip**

**AUSTRALIAN COSMETIC SURGERY**

November-January 2005  
Australian Cosmetic Surgery - pg 100  
VelaSmooth - "Cellulite & Fat Reduction"

**COSMETIC SURGERY MAGAZINE**  
INCORPORATING BODY HEALTH & BEAUTY

RRP \$10.95 (incl GST) (A2 01 198 incl GST)  
ISSUE 26 • NOV - JAN 2005

LIGHT THERAPY FOR ACNE